## Fabry-Perot observations at the 6m telescope

A.V. Moiseev, V.L. Afanasiev, and S.N. Dodonov

Special Astrophysical Observatory, Nizhnij Arkhyz, 369167, Russia

**Abstract.** A current status of the observations with Interferometer Fabry-Perot (IFP) at the 6m telescope are described. A reduction of the data obtaining with a new focal reducer SCORPIO in IFP mode are outlined.

## 1. SCORPIO design

The 2D spectroscopy observations with IFP at the 6m telescope using the CIGALE system were begun in early 80s by teams from Marseille Observatory (France) and from SAO (Boulesteix et al., 1982; Amram et al., 1992). In 1997 a CCD was attached on the old focal reducer instead of photon counter. A new multi-mode focal reducer SCORPIO (Spectral Camera with Optical Reducer for Photometrical and Interferometrical Observations) was developed in the SAO RAS in 2000 (Afanasiev et al., this Proceeding). Here we describe only Fabry-Perot mode of the SCORPIO. For more details see our WWW-site http://www.sao.ru/~moisav/scorpio.html.

The optics of SCORPIO compensates the aberrations of the main telescope mirror and provide a total focal ration F/2.9. The Queensgate IFPs ET-50 installs into the 35mm collimated beam. SCORPIO are under full remote control from a PC. A total DQE (telescope+filter+IFP+CCD) is about 20% in the  $H_{\alpha}$ spectral region. SCORPIO equipped with narrow filters (FWHM = 10 - 15 Å) for IFP observations in the  $H_{\alpha}$  line for object's velocities -200...+10.000 km s<sup>-1</sup>.

## 2. Calibration and data reduction

Focal reducer mounted on the prime focus universal adapter which has two movable probes for offset guiding and continuum and spectral lamps for calibration. A He-Ne-Ar lamp and continuum spectrum lamp mounted in the integration sphere which produced a uniform lighting. The optics of calibration system forms the F/4 beam on the reducer entrance. There is no systematic shift in velocities because telecentrical beams used for wavelength calibration.

A CCD has a large advantage in quantum effeciency in comparison with a photon counter. However a readout noise and a reading time do not allow to do very short exposures of the object channels. Therefore exist a problem of the photometry correction of flux and seeing in the individual channels because only 1-2 scanning cycles per channel may be accepted on CCD observations, we should take in account a time dependence of a sky brightness and of an atmospheric Moiseev et al.



Figure 1. SCORPIO observations of NGC 6951 in  $H_{\alpha}$ : raw IFP channel with night sky lines (**a**) and after night sky model removing (**b**). The wavelength cube channel were obtained from cube with sky lines (**c**) and with night sky lines removed – **d**.

extinction. Therefore wavelength object cube contains a set of narrow rings (Fig. 1c). This pattern limits the velocities measurements for the low brightness objects. We suggest a method to resolve this problem. Main idea is removing the sky counts *before* cube linerization. After standard CCD-reduction a mean radial profile of the sky lines is constructed by the azimuthal averaging of the interferometrical rings (Fig. 1a). The sky radial profile are subtracted from the object. Fig. 1b shows example of the sky subtracted images. Variations of an atmospheric extinction, seeing and channel offset shifts could be measured from the photometry of background stars and corrected in the cube. Unfortunately our method requires that about half of field of view will be free from the object emission, but it allows to obtain a good results in the object cube.

## References

- Amram P., Marcelin M., Bonnarel F., Boulesteix J., Afanas'ev V.L., Dodonov S.N., 1992, A&A, 263, 69
- Boulesteix J., Georgelin Y., Marcelin M., Fort J.A., 1982, in "Instrumentation for astronomy with large optical telescopes", 223